


Set de Agujas de Fístula A.V.





Set de Agujas de Fístula A.V.

Aguja de ala giratoria

Ala giratoria para una fácil manipulación.

Diseño mejorado del poro de la aguja

Desviación y mejora del flujo sanguíneo en múltiples direcciones.



Medidor	Longitud de aguja	Modelo	Tipo	Color de ala	Notas
17G	25mm	1.4×25GD/GS	Ala fija	Naranja	Con el ojo trasero, pinza, longitud del tubo de 300 mm, ala fija o giratoria disponible según su demanda.
		1.4×25XD/XS	Ala giratoria		
	32mm	1.4×32GD/GS	Ala fija		
		1.4×32XD/XS	Ala giratoria		
16G	25mm	1.6×25GD/GS	Ala fija	Verde	
		1.6×25XD/XS	Ala giratoria		
	32mm	1.6×32GD/GS	Ala fija		
		1.6×32XD/XS	Ala giratoria		
15G	25mm	1.8×25GD/GS	Ala fija	Azul	
		1.8×25XD/XS	Ala giratoria		
	32mm	1.8×32GD/GS	Ala fija		
		1.8×32XD/XS	Ala giratoria		

Esterilización: ETO.

* GD/XD: Aguja individual; GS/XS: Agujas gemelas
480 piezas/cartón para agujas gemelas 288
piezas/cartón para aguja individual.

Sede Corporativa

No.20 XingshanRoad, Weihai Torch Hi-tech Science Park, 264210 Weihai, Shandong Province, PEOPLE'S REPUBLIC OF CHINA.
Tel: +86 631 5660598 Fax: +86 631 5660598
www.wego-healthcare.com
PN:FNS20200702

Sede Negocios de Internacional

WEGO HEALTHCARE (SHENZHEN) CO., LTD.

25/F, East Technology Mansion, No.16, KeyuanRoad, Nanshan District, Shenzhen, China
Tel: +86 755 33892500 Fax: +86 755 33892508
E-mail: info@wego-healthcare.com



Certificate

No. Q8 094273 0012 Rev. 03

Holder of Certificate: **Shandong Weigao Group Medical
Polymer Co., Ltd.**
No.1 Weigao Road
Torch Hi-tech Science Park
264210 Weihai, Shandong Province
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Scope of Certificate: **The Provision of Ethylene Oxide and Radiation
Sterilization Services for Single Use Medical
Devices.**

The Certification Body of TÜV SÜD Product Service GmbH certifies that the company mentioned above has established and is maintaining a quality management system, which meets the requirements of the listed standard(s). All applicable requirements of the Testing, Certification, Validation and Verification Regulations TÜV SÜD Group have to be complied with. For details and certificate validity see: www.tuvsud.com/ps-cert?q=cert:Q8_094273_0012_Rev._03

Report No.: BJ24088401

Valid from: 2024-09-19
Valid until: 2026-09-30

Date, 2024-09-19



Christoph Dicks
Head of Certification/Notified Body

Certificate

No. Q8 094273 0012 Rev. 03

Applied Standard(s): ISO 13485:2016
(EN ISO 13485:2016/AC:2018, EN ISO 13485:2016/A11:2021)
Medical devices - Quality management systems -
Requirements for regulatory purposes

Facility(ies): **Shandong Weigao Group Medical Polymer Co., Ltd.**
No.1-1 Weigao Road, Torch Hi-tech Science Park, 264210 Weihai,
Shandong Province, PEOPLE'S REPUBLIC OF CHINA

The Provision of Ethylene Oxide and Radiation Sterilization
Services for Single Use Medical Devices.

Shandong Weigao Group Medical Polymer Co., Ltd.
No.1 Weigao Road, Torch Hi-tech Science Park, 264210 Weihai,
Shandong Province, PEOPLE'S REPUBLIC OF CHINA

The Provision of Management of Sterilization Services for Single
Use Medical Devices.